

July 17, 2009

SnugZ USA
5200 S Highland Dr
Ste 301
Salt Lake City, UT 84117

Attn: Barton Burdette

Job No: 115919A

LO

LABORATORY REPORT

Samples Received: One (1) Sample
Date Received: 07/06/2009
Purchase Order No: 14769

The sample was analyzed as follows:

Analysis


Page

Cadmium and Lead by SOP 7040, Rev 10
Phthalates by SOP 5290 Rev 1

2
3



Samina N. Hussain
Senior Chemist



Herman van Halbeek, Ph.D.
Senior Staff Chemist

This report is to be reproduced in its entirety.

Page 1 of 3

Cadmium and Lead by SOP 7040, Rev 10
Quantitative Analysis Report
Inductively Coupled Plasma-Mass Spectrometry

Parts Per Million (µg/g)

<u>Sample ID</u>	<u>Cadmium</u>	<u>Lead</u>
Power Cord	ND	0.2
Detection Limit:	0.1	0.1

Date Analyzed: 07-13-09

Quality Control Summary

Sample: Batch QC

<u>Analyte</u>	<u>Sample Result</u>	<u>Duplicate Result</u>	<u>Average Result</u>	<u>Sample RPD</u>	<u>Spike Conc</u>	<u>Spike Result</u>	<u>Spike % Rec</u>
Cadmium	ND	ND	ND	NA	299	324	108
Lead	0.30	0.30	0.30	NA	299	340	114

Date Analyzed: 07-13-09

Sample: Laboratory Fortified Blank (LFB)

<u>Analyte</u>	<u>Sample Result</u>	<u>Spike Conc</u>	<u>Spike Result</u>	<u>Spike % Rec</u>
Cadmium	ND	300	309	103
Lead	ND	300	311	104

Date Analyzed: 07-13-09

Phthalates by SOP 5290 Rev 1

Sample Preparation

A portion of each sample was extracted in methylene chloride. An aliquot of each extract was spiked with internal standards and was analyzed by GCMS.

GCMS Conditions

Column: 30 m x 0.32 mm Rtx-5Sil MS, 1 micron film
 Column Temp: 120 °C (hold 1 min) to 320 °C at 15 °C/min (hold 8 min)
 Injector Temp: 320 °C
 Mass Range: 41-475 amu; 3.35 scans/sec

Weight Percent (w/w)

Analyte	Powercord	Detection Limit
Dimethyl phthalate	ND	0.004
Diethyl phthalate	ND	0.004
Di-n-butyl phthalate	ND	0.004
Di-n-hexyl phthalate	ND	0.004
Butyl benzyl phthalate	ND	0.004
Bis(2-ethylhexyl)phthalate	ND	0.004
Di-n-octyl phthalate	ND	0.004
Diisononyl phthalate	ND	0.1
Diisodecyl phthalate	ND	0.1

Date analyzed: 07-08-09

Quality Control Summary

Sample ID:	Powercord						
Analyte	Sample Result	Spike Conc	Spike Result	Spike % Rec	Spike Duplicate Result	Spike Duplicate % Rec	Spike RPD
Dimethyl phthalate	ND	0.0391	0.0388	99	0.0410	105	6
Diethyl phthalate	ND	0.0397	0.0399	101	0.0423	107	6
Di-n-butyl phthalate	ND	0.0418	0.0452	108	0.0473	113	5
Di-n-hexyl phthalate	ND	0.0393	0.0461	117	0.0479	122	4
Butyl benzyl phthalate	ND	0.0390	0.0410	105	0.0426	109	4
Bis(2-ethylhexyl)phthalate	ND	0.0449	0.0525	117	0.0549	122	4
Di-n-octyl phthalate	ND	0.0416	0.0503	121	0.0519	125	3

QC Guidelines

Analyte	% Recovery	RPD Limit
All compounds	50 - 150	NMT 25